快速操作手册

1. 开启仪器电源和空气压缩机开关, 双击软件



2. 新增一个New Project



Project name:	Bioptic	
Result folder:	C.'Users/CQ43%Q-Analyzer	Brow

* 请输入项目名称以进入操作画面

3. 点选"Connect" 建立系统联机



卡夹产品列表:

卡夹 Alignment maker Mineral oil **Dilution buffer Separation** buffer 缓冲液槽 滴管 Alignment maker用离心管

~

Disconnect

Park

4.缓冲液与Marker的配置:

4-1.加缓冲液 标记"S"溶液槽加入Separation buffer 标记"C、W、P"三个凹槽中加入蒸馏水 *液面高度需到达指示线的位置

4-2.加Marker 吸取Alignment marker及Size marker各30µl加 入Marker离心管中,再加入30ul矿物油 (Mineral Oil) 在Marker上层

5.点选'Change buffer', 放入Marker和缓冲液

MA1: 放置Alignment Marker C109100; MB1: 放置Alignment Marker C109102; MA2: 放置Size Marker C109200; MB2: 放置Size Marker C109300:

缓冲液: 将缓冲液槽妥善放置于接盘处 *确定缓冲液槽放置方向正确

Park Wash Separation Clean

7. 点选"Change Sample" 打开透明门盖并放入样品(样品体积≥10µl) *请确定样本溶液中没有气泡

8. 新卡夹需先扎孔,若为使用过的卡夹,则无需重复扎孔,

9. 开启卡夹门, 放入卡夹, 再关紧卡夹门 *卡夹门之开关皆以按压方式作

10. 点选"Latch" 左下角信息方块会显示适配卡夹相关信息

快速操作手册

12-1. 点选"Sample Position", 选取样本位置并编辑样本信息 Edit

	Method Selector	and Andrew		×
it	Application	C RNA C RNA	A C Glycan	C Protein
鮮 📕	Alignment Marker	□ MA-1 ▼ 20	1000 C Reduce	e 🐨 Normal 🔿 Enhance
适	Cartridge Type	S1 💌 High resolu	ition cartridge(Shelf Life	: 6 Months)
	Method	Description	Range	Remark
	M-4-10-06-300	Sample injection 4kv 10s Separation 6kv 300s	15~1000 bp Best resolution: 2~4 bp	*
	M-4-10-06-500	Sample injection 4kv 10s Separation 6kv 500s	15~15k bp Best resolution: 2~4 bp	
	M-4-10-08-240	Sample injection 4kv 10s Separation 8kv 240s	15~5000 bp Best resolution: 4~10 bp	
,注 🗧	M-4-10-10-150	Sample injection 4kv 10s Separation 10kv 150s	15~5000 bp Best resolution: 10~50 bp	
法由	o M-8-10-06-300	Sample injection 8kv 10s Separation 6kv 300s	15~1000 bp Best resolution: 2~4 bp	For low concentration
小巴	M-8-10-06-500	Sample injection 8kv 10s Separation 6kv 500s	15~15k bp Best resolution: 2-4 bp	For low concentration sample
e 町	M-8-10-08-240	Sample injection 8kv 10s	155000 bp	For low concentration
	 High Voltage F 	Purge C Purge	Purge Modification	n
	Customized Me	thod		ок

*样本预期片段大小必须在所选择Alignment Marker范围内 (1) 如果 预期片段<1000bp, 选用"C109100"Alignment Marker (MA1) 作为内参,选用"C109200"Size Marker (MA2)

(2) 如果是 5000bp>预期片段≥1000bp,选用"C109102"Alignment Marker (MB1) 作为内参,选用"C109300"Size Marker (MB2)

12-3.点洗"Result Name"并输入结果文件名

		Sequen	ce	Open	Sav	/e Save as		_
SN	Sample	Method	Sample	Runs	Separation	Result Name		Add
	Position		Duration		Duration			Insert
1	A-01	M-4-10-06-300			300	For testing Sample ID	4	Delete

12-4.点选 "4"位置,勾选"Calculate",然后勾选"Create Size Marker", 'size marker injection time' 选择'10 S'. 点击'OK'确认

13.点选"Run"并启动实验

注意事项:

(1)确认Marker所放位置正确并且足量; (2)确认缓冲液槽内液面在指示线附近; (3)确认新卡夹已扎孔; (4)检测样本前需进行卡夹校正"Calibrate"。

下图为讯号图谱画面简介,详细说明请参阅使用手册

